

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2005-10	LOT # :	N0548816
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	<u>CaCl₂</u>	27	44	seconds	38	PASS
	<u>APC/CaCl₂</u>	--	92	seconds	56	PASS
APC Ratio		--	2.1	--	1.5	PASS

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	27	40	seconds	40	PASS
APC Ratio		1.3	1.8	--	1.6	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0548816	2005-10	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: D. Tencza

DATE: 05/20/04

APPROVAL SIGNATURE: J. Facciola

DATE: 05/20/04