CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2366, 25 mg Substrate	PRODUCT # :	82109039
EXP. DATE:	2008-01	LOT # :	N0152994
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	821090

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.6	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.04	PASS
RESIDUAL MOISTURE	0	5	%	0.15	PASS

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
S-2366	N0152994	2008-01	2 ml

PREPARED BY: V NIEVES DATE: 19/01/2005

APPROVAL SIGNATURE: R BUSTAMANTE DATE: 19/01/2005

QCM 821090, Rev. 04/00