

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2366, 25 mg Substrate	PRODUCT # : 82109039
EXP. DATE: 02/05	LOT # : N0225030
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 821090

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.07	PASS
RESIDUAL MOISTURE	0	5	%	0.5	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2366	N0225030	02/05	2 ml

PREPARED BY: Donna Tencza

DATE: 03/06/02

APPROVAL SIGNATURE: M. Fader

DATE: 03/06/02