CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST Factor VIII:C/4 Kit	PRODUCT # :	82191863
EXP. DATE:	2006-09	LOT # :	N0549093
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	821918

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
		MIN	MAX	UNITS		PASS/FAIL
Coatest Factor VIII/C4 Kit (82191863)						
High Range						
	Slope	0.0071	0.0106	--	0.0100	PASS
	Intercept	--	0.200	--	0.172	PASS
	Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range						
	Slope	0.026	0.051	--	0.043	PASS
	Intercept	--	0.300	--	0.194	PASS
	Standard Curve (r)	0.99	--	--	0.996	PASS

We hereby certify that the above product(s):
- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP.DATE	VOLUME
$CaCl_2$	N0347875	2007-03	6 ml (x 1)
S-2222/ I-2581	N0348102	2006-09	6 ml (x 1)
Factor Reagent	N0548913	2006-11	2 ml (x 4)
Buffer, pH 7.3	N0348138	2006-09	23 ml (x 1)
Phospholipid	N0448322	2006-09	2.3 ml (x 1)

PREPARED BY: *Raquel Bustamante* DATE: 06/04/04

APPROVAL SIGNATURE: *John Facciola* DATE: 06/04/04

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000