

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: Control Plasma-LMW Heparin	PRODUCT # : 823492
EXP. DATE: 2007-02	LOT # : N0940623
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 823500/823492

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
A. Control Plasma, Low: LMW Heparin (H01337)					
Mean value	0.33	0.42	IU/vial	0.34	PASS
Intra series CV (n=60)	-	10	%	9.1	PASS
Inter bottle CV (n=12)	-	10	%	7.7	PASS
Total % CV (n=12)	-	7	%	6.1	PASS

B. Control Plasma, High, LMW Heparin (H01346)					
Mean value	0.64	0.84	IU/vial	0.69	PASS
Intra series CV (n=60)	-	5	%	3.8	PASS
Inter bottle CV (n=12)	-	5	%	3.6	PASS
Total % CV (n=12)	-	7	%	3.4	PASS

C. Control Plasma LMW Heparin Kit					
Slope	-1.063	-0.523	-	-0.705	PASS
r ²	0.98	-	-	0.997	PASS
Control (low) recovery	0.27	0.41	IU/mL	0.29	PASS
Control (high) recovery	0.62	0.76	IU/mL	0.71	PASS

Kit Components:

Name	Lot Number	Exp. Date	Mfg. Date
Control Plasma, Low, LMW Heparin	N0840513	2007-02	8/16/2004
Control Plasma, High, LMW Heparin	N0840514	2007-02	8/16/2004

THE PLASMA MATERIAL USED IN THE MANUFACTURE OF THIS PRODUCT HAS BEEN TESTED AND FOUND NON-REACTIVE FOR HIV, HBsAg AND ANTI-HCV.

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: NPTAEL

DATE: 8/27/2004

APPROVAL SIGNATURE: JOHN FACCIOLA

DATE: 8/27/2004