CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	CALIBRATION PLASMA	PRODUCT # :	82353463
EXP. DATE:	2007-02	LOT # :	N0347839
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	823534

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
Clarity (A_{405})	--	0.400	--	0.356	PASS
CHROMOGENIC					
Antithrombin Activity	80	130	%	103	PASS
Plasmin Inhibitor Activity	80	130	%	93	PASS
Plasminogen Activity	80	130	%	90	PASS
Protein C Activity	80	130	%	103	PASS
Factor VIII Activity	80	130	%	97	PASS
Free Protein S	80	120	%	106	PASS
von Willebrand Factor	85	140	%	102	PASS

Kit Components:

Name	Lot Number	Exp. Date	Volume
Calibration Plasma	N0347839	2007-02	1 ml

PREPARED BY: *D. Tenoza* DATE: 05/26/04

APPROVAL SIGNATURE: *J. Facciola* DATE: 05/26/04

Form QCM 823534, Rev. 10/02