

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	CALIBRATION PLASMA	PRODUCT # :	82353463
EXP. DATE:	2005-04	LOT # :	N0632955
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	823534

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
Clarity (A ₄₀₅)	--	0.400	--	0.393	PASS
<u>CHROMOGENIC</u>					
Antithrombin Activity	80	130	%	90	PASS
Plasmin Inhibitor Activity	80	130	%	95	PASS
Plasminogen Activity	80	130	%	88	PASS
Protein C Activity	80	130	%	87	PASS
Factor VIII Activity	80	130	%	111	PASS
Free Protein S	80	120	%	92	PASS
von Willebrand Factor	85	140	%	119	PASS

Kit Components:

Name	Lot Number	Exp. Date	Volume
Calibration Plasma	N0632955	2005-04	1 ml

PREPARED BY: Donna Teneza DATE: 08/19/03

APPROVAL SIGNATURE: J. Facciola DATE: 08/19/03