CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	CALIBRATION PLASMA	PRODUCT # :	82353463
EXP. DATE:	2007-08	LOT # :	N0940810
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	823534

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
Clarity (A_{405})	--	0.400	--	0.340	PASS
CHROMOGENIC					
Antithrombin Activity	80	130	%	100	PASS
Plasmin Inhibitor Activity	80	130	%	91	PASS
Plasminogen Activity	80	130	%	85	PASS
Protein C Activity	80	130	%	95	PASS
Factor VIII Activity	80	130	%	81	PASS
Free Protein S	80	120	%	93	PASS
von Willebrand Factor	85	140	%	99	PASS

Kit Components:

Name	Lot Number	Exp. Date	Volume
Calibration Plasma	N0940810	2007-08	1 ml

We hereby certify that the above product(s):
-have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
-meet all applicable requirments of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: *D. Tencza* DATE: 12/02/04

APPROVAL SIGNATURE: *J. Facciola* DATE: 12/02/04

Form QCM 823534, Rev. 05/04