CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	CONTROL PLASMA NORMAL	PRODUCT # :	82354263
EXP. DATE:	2007-05	LOT # :	N0649325
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	823542

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
Clarity (A_{405})	--	0.400	--	0.381	PASS
CHROMOGENIC					
Antithrombin Activity	80	--	%	104	PASS
Plasmin Inhibitor Activity	80	--	%	100	PASS
Plasminogen Activity	80	--	%	88	PASS
Protein C Activity	80	--	%	102	PASS
Factor VIII Activity	80	--	%	95	PASS
Free Protein S	80	120	%	97	PASS

Kit Components:

Name	Lot Number	Exp. Date	Volume
Normal Control Plasma	N0649325	2007-05	1 ml

We hereby certify that the above product(s):
- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: *D. Tencza* DATE: 09/15/04

APPROVAL SIGNATURE: *J. Facciola* DATE: 09/15/04

Form QCM 82354263, Rev. 11/03